

Sichuan Testing Center of Medical Devices
Notification of testing result



Report No. QX2020H03808

SampleName	STERILE FACE MASK	Model	\
Requester	Chengdu Xinjin Shifeng Medical Apparatus & Instrument Co.,Ltd	Batch/lot No.	Lot 200330
		Date of Manufacture	20200330
Inspector	\	Factory Serial Number	\
Manufacturer	Chengdu Xinjin Shifeng Medical Apparatus & Instrument Co.,Ltd	Packing	\
Trade mark	\	Sample status	\
Test category	Contract	Sampling location	\
SamplingMethod	\	Sample quantity	30
SampleSize	\	Sent / Sampled by	\
Date of Receipt	2020-04-01	Test Location	Sichuan Test Center of Medical Devices
Expiration Date	20220329	TestDate	2020-04-01 to 2020-04-14
Test Standard	EN 14683:2019+AC:2019Medical face masks - Requirements and test methods		
Test Item	Partial test		
Test Result	This product is inspected under Reference of EN 14683:2019+AC:2019Medical face masks - Requirements and test methods, Testing results are available in attached pages.		
Note	\		



Tested by: 朱蕊 Reviewed by: 刘延军 Approved by: 刘兴兰
Approval date: 2020-04-14

Attachment of testing result

Report No.QX2020H03808

SampleName	STERILE FACE MASK	Sample quantity	30
Manufacturer	Chengdu Xinjin Shifeng Medical Apparatus & Instrument Co.,Ltd	Batch/lot No	Lot 200330

Test result		
Test Item	Requirements (Type II)	Result
5.2.2Bacterial Filtration Efficiency (BFE)	$\geq 98\%$	Minimum value: 99.7%
5.2.3Breathability(Differential Pressure)	$< 40\text{Pa}/\text{cm}^2$	$30\text{Pa}/\text{cm}^2 \sim 36\text{Pa}/\text{cm}^2$

Remarks: 5.2.2Bacterial Filtration Efficiency (BFE) The side of the test specimen was facing towards the challenge aerosol: out side of mask

成本
 专用